

MUNICIPAL CORPORATION OF DELHI  
KASTURBA HOSPITAL  
DARYA GANJ, NEW DELHI-110002

KH/Purchase/2022-23/12

Dated: 14/03/2023

Subject: Invitation for Quotations.

Quotations are invited for the Purchase of Electric Tissue Floatation Water Bath for Hispathology Lab, Pathology Department Kasturba Hospital.

S.No	Description of work/Item As Below	Total no. of Items. As below	Quotation to be submitted in the Office of Addl. M.S./KH Latest by 2:00 pm on 24.03.2023 and same will be opened on same day at 2:30PM
1.	Electric Tissue Floatation Water Bath for Histopathology Lab	02	

**Technical Specifications:-**

1. Electric Tissue floatation water bath for processing of histopathology tissue sections.
2. It should have temperature control thermostat with working temp. from ambient to 70 deg. C with temperature adjustability.
3. Set temperature tolerance of +/- 2 deg. C, Temperature to be controlled by capillary type thermostat.
4. It should be double walled, single piece comprising of outer & inner chambers.
5. Outer body of aluminum with powder coating, inner vessel of black anodized aluminum.
6. It should have embedded heating element plate placed between inner and outer chambers and completely separated from water contact.
7. It should have high grade insulation, filled around.
8. Inner vessel of minimum 2 liters capacity - removable for easy cleaning.
9. Inner chamber diameter to be between 22-26 cm.
10. Display should show on/off(power indicator), set temperature & actual temp.
11. It should have wide rim enough for vetting the wet slides.
12. Should provide with power cord of min. 3 metre length & appropriate plug.
13. Atleast 2 years' warranty to be provided.
14. It should be BIS/FDA/CE Certified; it should be ISI marked.
15. Personnel and spare parts for after sales service must be readily available in Delhi/NCR.

**The firms should submit the following:-**

- (a) Copy of PAN Number
- (b) Copy of GST Registration No.
- (c) Should provide at a rate which is lesser than or equal to at which supplied to govt. hospital including MCD.
- (d) Should have the minimum 5 years experience of eligibility criteria.
- (e) Should not be blacklisted.

  
Addl.MS  
Kasturba Hospital

- Distribution: (1) A0(IT) for uploading on MCD Website (through E-mail)  
(2) Notice Board Kasturba Hospital.  
(3) Notice Board of Hindu Rao Hospital, RBIPMT, GLMH, SDNH, ID Hospital, GB Pant and LNJP Hospital (through E-Mail)  
(4) HOD (Pathology) /KH  
(5) Office Copy