

MUNICIPAL CORPORATION OF DELHI
KASTURBA HOSPITAL
DARYA GANJ, NEW DELHI-110002

KH/Purchase/2022-23/09

Dated: 09/04/2023

Subject: Invitation for Quotations.

Quotations are invited for the Purchase of Hot Air Oven for NICU, Pediatric Department, Kasturba Hospital.

S.No	Description of work/Item As Below	Total no. of Items. As below	Quotation to be submitted in the Office of Addl. M.S./KH Latest by 2:00 pm on 20.02.2023 and same will be opened on same day at 2:30PM
1.	Hot Air Oven	01	

Technical Specifications:-

1. It should have capacity of 125 liters.
2. Power should be 1500-2000 watts.
3. Ambient temperature should be 250 °C with temperature accuracy $\pm 1^{\circ}$ C.
4. Insulation should be of 75mm Mineral wool.
5. It should have microprocessor based PID control.
6. It should have LED display. It should display Time, Temperature settings and Temperature reading.
7. It should have 3 removable Aluminum Perforated shelves.
8. Inner chamber should be of Stainless steel (304grade).
9. Outer chamber should be of MS with powder coated paint.
10. Door should be fitted with heavy hinges and should be insulated and should have locking facility.
11. There should be Motorized blower of 0.5HP for forced air circulation.
12. Blower motor should be totally enclosed and should be air cooled.
13. Length of power cord should be approx 3 meter with compatible Indian plug.
14. It should be operable on temperature of 10 to 40 degree Celsius and 15-90% humidity.
15. It should have over temperature protection.
16. It should have safety alarm system.
17. Oven should be corrosion resistant.
18. Instructions manual should be provided.
19. It should operate on power supply of 220 volts, 50Hz single phase.
20. Approximate weight should be 30 kg.
21. It should be BIS/Indian Standards/CE certified.
22. It should be ISO certified.

GENERAL TERMS & CONDITIONS

- Firm Should have wide channel partner network, factory trained customer support personnel available in Delhi with adequate stock of spares & facility for onsite repair of the system to be provided.
- Firm should ensure and certify the availability of spare parts for 10 yrs from the date of commissioning /installation.
- Warranty to be given for 5 years.
- Firm should provide an undertaking that they will provide CMC for next 5 years after expiry of warranty period.
- Rates for CMC along with consumables after period of warranty to be provide
- Demonstration of equipment in Delhi or NCR, in working condition is a must at the cost of vendor/supplier.
- List of users along with performance report of the quoted /similar model from Govt Hospital/private hospital of repute to be provided
- The equipment is to be provided complete with all the accessories so that it can be used without any further requirement of accessories.
- Firm should be preferably manufacturer /authorized agent of principle with latest certificate for this particular tender.

- Firm should give an undertaking that they have not supplied the machine at rates lower than this to some other hospital.
- Firm should give a certificate that it is not blacklisted or no vigilance /CBI enquiry is pending against it.
- The firm should be reputed one and quote with good, latest quality equipment.

The firms should submit the following:-

- (a) Copy of PAN Number
- (b) Copy of GST Registration No.
- (c) Should provide at a rate which is lesser than or equal to at which supplied to govt. hospital including MCD.
- (d) Should have the minimum 5 years experience of eligibility criteria.
- (e) Should not be blacklisted.


Addl.MS
Kasturba Hospital

- Distribution: (1) A0(IT) for uploading on MCD Website .
(2) Notice Board Kasturba Hospital.
(3) Notice Board of Hindu Rao Hospital, RBIPMT, GLMH,
SDNH, ID Hospital, GB Pant and LNJP Hospital (through E-Mail)
(4) HOD (Pediatric) /KH
(5) Office Copy